

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20201005
 Production date 05 October 2020
 Expiry date 31 October 2025

Article No.
 IC-5-30°-s
 IC-5-35°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 18 November 2020
 Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann