

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20201209
 Production date 09 December 2020
 Expiry date 31 December 2025

Article No.

Bp-20-35°-bl

| | Results | Specifications |
|--|----------------|-----------------------|
| Packaging | PASSED | PASSED |
| Sterilization | PASSED | PASSED |
| Mouse Embryo Assay (1-cell, blastocysts after 96 hours) | 100 | ≥ 80 % |

Release date 20 January 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann