

**Certificate of Analysis  
Micropipettes - ICSI**

Batchnumber	20201209
Production date	09 December 2020
Expiry date	31 December 2025
<b>Article No.</b>	
IC-5-30°-s	

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	20 January 2021
Released by Signature	Dr. Julia Heinzmann

*p.p. Julia Heinzmann*

REV 04\_00 / 2020-01-27