

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber		20210129
Production date		29 January 2021
Expiry date		31 January 2026
	Article No.	
	IC-5-35°-s	

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	01 March 2021
Released by Signature	Dr. Julia Heinzmann

p.p. Julia Heinzmann