

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber	20210215
Production date	15 February 2021
Expiry date	28 February 2026
Article No.	
IC-5-30°-s	

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	29 March 2021
Released by Signature	Dr. Julia Heinzmann

p.p. Julia Heinzmann