

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20210325
 Production date 25 March 2021
 Expiry date 31 March 2026

Article No.
 IC-5-30°s
 IC-5-35°s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 12 May 2021
 Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann