

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20210422
 Production date 22 April 2021
 Expiry date 30 April 2026

Article No.
 IC-5-30°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 15 June 2021

Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann