

**Certificate of Analysis  
Micropipettes - Holding**

Batchnumber 20210628  
 Production date 28 June 2021  
 Expiry date 30 September 2026

**Article No.**

HP-120-30°  
 HP-100-35°

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 04 October 2021

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*