

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20210805
 Production date 05 August 2021
 Expiry date 31 August 2026

Article No.
 IC-5-30°-s
 IC-5-35°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 14 October 2021
 Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann