

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20210910
 Production date 10 September 2021
 Expiry date 30 September 2026

Article No.

HP-100-30°
 HP-110-35°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 01 November 2021
 Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann