

Certificate of Analysis Micropipettes - Holding

Batchnumber	20211005
Production date	05 October 2021
Expiry date	31 October 2026

Article No.

HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date	18 November 2021
--------------	------------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann

REV 03_00 / 2020-01-27