

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20211019
 Production date 19 October 2021
 Expiry date 31 October 2026

Article No.

BP-30-30°-bl

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 30 November 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann