

**Certificate of Analysis  
Micropipettes - Biopsy**

Batchnumber 20211124  
 Production date 24 November 2021  
 Expiry date 30 November 2026

**Article No.**

BP-25-30°-bl

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 24 January 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*