

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20211124
 Production date 24 November 2021
 Expiry date 30 November 2026

Article No.

HP-110-30°
 HP-120-30°

| | Results | Specifications |
|--|----------------|-----------------------|
| Packaging | PASSED | PASSED |
| Sterilization | PASSED | PASSED |
| Mouse Embryo Assay (1-cell, blastocysts after 96 hours) | 90 | ≥ 80 % |

Release date 24 January 2022
 Released by Dr. Julia Heinzmann
 Signature

p.p. Julia Heinzmann