

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20211124
 Production date 24 November 2021
 Expiry date 30 November 2026

Article No.
 IC-5-30°-s
 IC-7-35°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 24 January 2022
 Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann