

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20220127
 Production date 27 January 2022
 Expiry date 31 January 2027

Article No.

HP-100-30°
 HP-110-30°
 HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 30 March 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann