

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20220207
 Production date 07 February 2022
 Expiry date 28 February 2027

Article No.
 IC-5-35°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date 06 April 2022
 Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann