

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20220221
 Production date 21 February 2022
 Expiry date 28 February 2027

Article No.
 BP-25-30°-bl

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date 06 April 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann