

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20220304
 Production date 04 March 2022
 Expiry date 31 March 2027

Article No.

HP-100-30°
 HP-100-35°
 HP-110-35°
 HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 09 May 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann