

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20220329
 Production date 29 March 2022
 Expiry date 31 March 2027

Article No.

BP-20-35°-bl
 BP-30-30°-bl

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 09 May 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann