

**Certificate of Analysis  
Micropipettes - ICSI**

Batchnumber 20220329  
 Production date 29 March 2022  
 Expiry date 31 March 2027

**Article No.**  
 IC-5-35°-s

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 09 May 2022  
 Released by Signature Dr. Julia Heinzmann

*p.p. Julia Heinzmann*