

**Certificate of Analysis  
Micropipettes - Holding**

Batchnumber 20220502  
 Production date 31 May 2022  
 Expiry date 31 May 2027

**Article No.**

HP-100-30°  
 HP-120-30°

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 04 August 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*