

**Certificate of Analysis  
Micropipettes - Biopsy**

Batchnumber	20220623
Production date	23 June 2022
Expiry date	30 June 2027

**Article No.**

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 04 August 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*