

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20220623
 Production date 23 June 2022
 Expiry date 30 June 2027

Article No.
IC-5-30°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 04 August 2022
 Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann