

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20220902
 Production date 02 September 2022
 Expiry date 31 May 2027

Article No.

HP-110-30°
 HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 26 October 2022

Released by Vanessa Bierlich

Signature

REV 03_00 / 2020-01-27