

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20220902
 Production date 02 September 2022
 Expiry date 31 May 2027

Article No.
 IC-5-35°-s
 IC-SL-4.6-30°-s
 IC-SL-4.6-35°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 26 October 2022
 Released by Signature Vanessa Bierlich

P.P. 

REV 04_00 / 2020-01-27