

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20220905
 Production date 05 September 2022
 Expiry date 31 May 2027

Article No.
 HP-100-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 15 November 2022
 Released by Nicole Geiger

Signature

