

**Certificate of Analysis
Micropipettes - ICSI**


Batchnumber 20220923
 Production date 30 September 2022
 Expiry date 30 September 2027

Article No.

IC-5-30°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 05 January 2023
 Released by Signature Nicole Geiger



REV 04_00 / 2020-01-27