

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20220926
 Production date 30 September 2022
 Expiry date 30 September 2027

Article No.


HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 05 January 2023

Released by Nicole Geiger

Signature



REV 03_00 / 2020-01-27

