

**Certificate of Analysis  
Micropipettes - ICSI**


Batchnumber 20221014  
 Production date 31 October 2022  
 Expiry date 31 October 2027

**Article No.**

IC-5-35°-s  
 IC-SL-4.6-30°-s  
 IC-5-30°-s

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 05 January 2023  
 Released by Signature Nicole Geiger



REV 04\_00 / 2020-01-27

