

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20221215
 Production date 15 December 2022
 Expiry date 31 December 2027

Article No.
 IC-5-30°-s
 IC-7-35°-s

| | Results | Specifications |
|--|----------------|-----------------------|
| Packaging | PASSED | PASSED |
| Sterilization | PASSED | PASSED |
| Mouse Embryo Assay (1-cell, blastocysts after 96 hours) | 95 | ≥ 80 % |

Release date 31 January 2023

Released by Signature Nicole Geiger

