

Certificate of Analysis Micropipettes - Holding

Batchnumber	20230413
Production date	30 April 2023
Expiry date	30 April 2028

Article No.

HP-100-30°
HP-100-35°
HP-110-35°
HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	08 June 2023
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 03_00 / 2020-01-27