

Certificate of Analysis Micropipettes - Biopsy

Batchnumber	20230815
Production date	31 August 2023
Expiry date	31 August 2028

Article No.

BP-20-30°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	81	≥ 80 %

Release date	24 October 2023
--------------	-----------------

Released by	Nicole Geiger
-------------	---------------

Signature



REV 04_00 / 2023-09-15

