

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20230830
 Production date 31 August 2023
 Expiry date 31 August 2028

Article No.


BP-25-30°-bl
 BP-30-30°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	81	≥ 80 %

Release date 24 October 2023

Released by Nicole Geiger

Signature



REV 04_00 / 2023-09-15