

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20230912
 Production date 30 September 2023
 Expiry date 30 September 2028

Article No.

BP-25-30°-bl
 BP-30-30°-bl
 BP-30-35°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 14 November 2023

Released by Nicole Geiger

Signature



REV 04_00 / 2023-09-15