

**Certificate of Analysis
Micropipettes - Holding**


Batchnumber 20231114
 Production date 30 November 2023
 Expiry date 30 November 2028

Article No.

HP-110-30°
HP-110-35°

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay <small>(1-cell, blastocysts after 96 hours)</small>	100	≥ 80 %

Release date 23 January 2024
 Released by Nicole Geiger
 Signature



REV 04_00 / 2023-09-15