

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20240220
 Production date 28 February 2024
 Expiry date 28 February 2029


Article No.

HP-100-30°
 HP-120-30°

| Parameter | Results | Specifications |
|--|---------|----------------|
| Packaging | PASSED | PASSED |
| Sterilization | PASSED | PASSED |
| Mouse Embryo Assay (1-cell, blastocysts after 96 hours) | 100 | ≥ 80 % |

Release date 17 April 2024
 Released by Nicole Geiger

Signature



REV 04_00 / 2023-09-15