

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20240403
 Production date 30 April 2024
 Expiry date 30 April 2029

Article No.

IC-SL-4.6-30°-s

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	81	≥ 80 %

Release date 16 May 2024

Released by Nicole Geiger

Signature



REV 05_00 / 2023-09-15