

**Certificate of Analysis  
Micropipettes - Biopsy**


Batchnumber 20240422  
 Production date 22 April 2024  
 Expiry date 30 April 2029

**Article No.**

BP-20-30°-bl  
 BP-25-30°-bl  
 BP-30-35°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 28 June 2024  
 Released by Nicole Geiger  
 Signature



REV 04\_00 / 2023-09-15