

**Certificate of Analysis  
Micropipettes - Holding**

Batchnumber 20240722  
 Production date 22 July 2024  
 Expiry date 31 July 2029

**Article No.**


HP-100-35°

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 30 August 2024

Released by Nicole Geiger

Signature



REV 04\_00 / 2023-09-15