

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20240722
 Production date 22 July 2024
 Expiry date 31 July 2029

Article No.


IC-7-35°-be

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 30 August 2024

Released by Nicole Geiger

Signature



REV 05_00 / 2023-09-15