

**Certificate of Analysis  
Micropipettes - Holding**


Batchnumber 20240828  
 Production date 28 August 2024  
 Expiry date 31 August 2029

**Article No.**

HP-110-35°

<b>Parameter</b>	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date 02 October 2024  
 Released by Nicole Geiger  
 Signature



REV 04\_00 / 2023-09-15