

**Certificate of Analysis  
Micropipettes - Holding**

Batchnumber 20250131  
 Production date 31 January 2025  
 Expiry date 31 January 2030

**Article No.**


HP-120-30°

<b>Parameter</b>	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 04 March 2025

Released by Nicole Geiger

Signature



REV 04\_00 / 2023-09-15