

## Certificate of Analysis Micropipettes - Holding

**Batchnumber** 20250220

**Expiry date** 28 February 2030

**Article No.**

HP-100-30°


HP-120-30°

| Parameter                                                  | Results | Specifications |
|------------------------------------------------------------|---------|----------------|
| Packaging                                                  | PASSED  | PASSED         |
| Sterilization                                              | PASSED  | PASSED         |
| Mouse Embryo Assay<br>(1-cell, blastocysts after 96 hours) | 100     | ≥ 80 %         |

**Release date** 31 March 2025

**Released by** Nicole Geiger

**Signature**



REV 04\_00 / 2023-09-15