

Certificate of Analysis Micropipettes - ICSI

Batchnumber 20250220

Expiry date 28 February 2030

Article No.

IC-SL-4.6-30°-s

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 31 March 2025

Released by Nicole Geiger

Signature

pp. *N. Geiger*

REV 05_00 / 2023-09-15