

Certificate of Analysis Micropipettes - Holding

Batchnumber	20250425
Production date	25 April 2025
Expiry date	30 April 2030

Article No.

HP-100-30°

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date	03 June 2025
Released by	Nicole Geiger

Signature

pp. N. Geiger

REV 04_00 / 2023-09-15