

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20250425
Production date 25 April 2025
Expiry date 30 April 2030

Article No.

IC-5-30°-s

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 03 June 2025
Released by Nicole Geiger
Signature

pp. N. Geiger

REV 05_00 / 2023-09-15