

Certificate of Analysis Micropipettes - Biopsy

Batchnumber	20250623
Production date	23 June 2025
Expiry date	30 June 2030

Article No.

BP-35-30°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	30 July 2025
--------------	--------------

Released by	Nicole Geiger
-------------	---------------

Signature

pp. N. Geiger

REV 04_00 / 2023-09-15