

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20250821
Production date 21 August 2025
Expiry date 30 August 2030

Article No.

BP-30-30°-bl
BP-35-35°-be
BP-35-35°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 30 September 2025
Released by Nicole Geiger
Signature

pp. P. Geiger

REV 04_00 / 2023-09-15