

Certificate of Analysis Micropipettes - ICSI

Batchnumber	20250821
Production date	21 August 2025
Expiry date	31 August 2030

Article No.

IC-SL-4.6-30°-s

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	30 September 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 05_00 / 2023-09-15